

**ALARM****LIBERO NORMAL 2-30 °C EU****Additional Information**

Instructions:

Please follow instructions on the shipment inspection sheet that was enclosed with the shipment.

LIBERO Configuration

Type:	LIBERO CS V4.32	Current State:	Device not started
Device ID:	7105 0139 783	Start at:	Not available
Log Interval:	10 m	First Measurement:	Start Delay: 1 h 0 m
Report Time Base:	UTC	Checksum:	C230 / 1.051.777.117
Configured by:	S1770, EC60/mscheidbach, 10.Apr.2019 12:38:43		

Alarm Zone	Allowed Time	Total Time	No. of Excursions	Longest Exc.	Status
H2: over 40.0 °C	0 m (sin)	0 m	0 / unlimited	0 m	OK
H1: over 30.0 °C	3 d 0 h 0 m (cum)	0 m	0 / unlimited	0 m	OK
G: 2.0 to 30.0 °C	unlimited	2 d 7 h 0 m			
L1: below 2.0 °C	14 d 0 h 0 m (cum)	0 m	0 / unlimited	0 m	OK
L2: below -20.0 °C	0 m (sin)	0 m	0 / unlimited	0 m	OK

MKT Limits and Maximum Duration

Duration: 2 d 7 h 0 m
MKT: 20.5 °C

Alarm Conditions

none

Logging Results

Highest Temperature:	24.4 °C; 24.May.2019 18:53:14	Stop at:	27.May.2019 08:33:30
Lowest Temperature:	18.8 °C; 25.May.2019 23:53:14	Alarm at:	11.Feb.2022 19:26:27
Average Temperature:	20.4 °C	File created:	11.Feb.2022 19:26:29

[°C] **WARNING: Device Error E003**